ORIGINAL ARTICLE

Characteristic Analysis and Design Optimization of Bubble Artificial Muscles

Richard Suphapol Diteesawat,^{1,2} Tim Helps,^{1,2} Majid Taghavi,^{1,2} and Jonathan Rossiter^{1,2}

Abstract

Soft robotics requires new actuators and artificial muscles that are lighter, less expensive, and more effective than current technologies. Recently developed bubble artificial muscles (BAMs) are lightweight, flexible, inexpensive, pneumatic actuators with the capability of being scalable, contracting at a low pressure, and generating sufficient tension and contraction for assisting human mobility. The BAMs are simply fabricated by using a commercial plastic tubing with retaining rings, forming a “bubble” shape and creating a series of contractile units to attain a desired stroke. They can deliver high contraction through optimization of actuator length and radius, or high tension by strengthening their materials to operate at high pressure. Here, we present a detailed analysis of BAMs, define a model for their actuation, and verify the model through a series of experiments with fabricated BAM actuators. In tests, a maximum contraction of 43.1% and a maximum stress of 0.894 MPa were achieved, corresponding to the BAM lifting a load 1000 times its own weight (5.39 g). The BAM model was built to predict experimental performance, for example, the relationship between tension and contraction at various applied pressures, and between contraction and pressure. Characteristic analysis and design optimization of the BAM are presented as an approach to design and manufacture the ideal “bubble” actuator at any required dimensions. A BAM orthosis is demonstrated as assisting a sit-to-stand transition on a leg mechanism, constructed to match the scale of a human’s lower limb. Guidelines for further improvement of the BAM are also included.

Keywords: wearable, actuator, pneumatic, fluidic, artificial muscle

Introduction

ASSISTIVE TECHNOLOGIES HAVE emerged to endow humans with more capabilities and independence in life, especially among older adults. Powered exoskeletons and orthoses have considerably developed over the past decades to provide wearers with more strength and mobility.¹ These rigid exoskeletons can be divided into two categories based on their purposes: (1) enabling and restoring mobility in patients with disabilities and (2) strengthening workers and soldiers to undertake heavy activities for industrial or military applications.² However, there are many ongoing challenges and requirements, including low weight, flexibility, comfort and adaptability to human body, esthetics, affordability, long lifetime, reliability, and safety. In addition, restriction of natural body motion and damage to the body of healthy individuals must be avoided.

Soft robots have advantages over conventional rigid robots thanks to their light weight and compliance, being more suitable for direct interactions with the human body while potentially reducing harm. Pneumatic artificial muscles (PAMs) are one of the most commonly used actuators to drive assistive soft exoskeletons.³ Although pneumatic exoskeletons were initially made of mostly rigid components,^{4,5} more recently they have been developed with more flexible, compliant elements.^{6–15} Soft exoskeletons and artificial muscles have also used a range of power sources, including cable-tendon-driven mechanisms, that is, Exosuit,¹⁶ Myosuit,¹⁷ and XoSoft¹⁸; direct electro-mechanical energy transduction in polymers, for example, polyvinyl chloride (PVC) gel,^{19,20} dielectric elastomers,^{21,22} dielectrophoretic liquid zipping actuators,²³ and Peano-HASEL²⁴; and thermo-mechanical actuation such as coiled polymer.²⁵

¹Department of Engineering Mathematics, University of Bristol, Bristol, United Kingdom.

²Bristol Robotics Laboratory, Bristol, United Kingdom.

PAMs are soft, flexible contractile actuators that change shape and contract when activated by pressurized air.²⁶ The McKibben pneumatic muscle is widely used in soft exoskeletons. It is made of a length of an elastic tubing enclosed by a braided sleeve, which contracts to form a cylindrical shape under an applied pressure.²⁷ Similarly, straight fiber PAMs use an elastic tube featuring reinforced fibers to increase membrane strength, and they form a spherical shape when actuated.^{28–31}

Other soft pneumatic actuators (SPAs) apply this concept of using a stiff embedded sheet on the inside of soft elastomer³² or creating a three-dimensional (3D)-printed soft contractile actuator consisting of stiff and soft composites.³³ These PAMs use an elastic material as the actuator's membrane, which stretches and expands under high applied pressure, and use a stiffer material to limit membrane expansion, resulting in a pre-determined shape and contraction.

On the other hand, pleated pneumatic artificial muscles (PPAMs) only use a high-strength inelastic material as an actuator membrane constrained by special end fittings to create equal radial pleats at the actuator ends, contracting to an elliptical shape under pressure.³⁴ The PPAMs were initially used as an artificial muscle for a gripper, a robot arm, and a walking or running mechanism.³⁵ Pouch motors use a commercial flexible plastic material to create a lightweight series contractile actuator, allowing low-pressure actuation.³⁶ Bending SPAs, made of a thin-walled flexible tubing, have achieved large bending motions and high torque at low pressure operation.³⁷ These three actuators benefit from their inelastic material's strength to operate under pressure and produce contraction and tensile force.

Recently, new contractile PAMs made of a commercial inelastic plastic tubing with the addition of soft rubber rings (series pneumatic artificial muscle—sPAM) or rigid metal rings (bubble artificial muscles—BAMs) have been presented. The sPAM³⁸ was developed for navigation and survey applications, enabling a low-pressure soft continuum robot to steer and operate in constrained and cluttered environments,³⁹ whereas the BAM was developed as a human-like muscle to assist human mobility, for example, for aiding knee flexion during walking.⁴⁰

The BAMs are lightweight, compliant, and inexpensive pneumatic muscles, designed to have similar structure and function as the PPAM while being considerably simpler, less expensive, and lower weight; being made of a thin, flexible inelastic tubing and stiff retaining rings. This grants the BAM flexibility and low- to high-pressure actuation to deliver either high contraction or high tensile force depending on the thickness and stiffness of the tubing material (Fig. 1; Supplementary Movie S1).

The BAM can be designed to deliver the most suitable mechanical assistance for different parts of the human body, which requires different tensile forces and amounts of contraction. To enable this design flexibility, we present characteristic analysis and design optimization of BAMs, develop an actuator model encompassing the unique buckled folds at the rings, and verify the model against experimentation with a range of BAM actuators.

Pneumatic Actuators

Pleated pneumatic artificial muscle

We first consider the PPAM³⁴ (Supplementary Fig. S1A), from which models of the sPAM and the BAM are derived.

FIG. 1. Actuation of BAMs. (A) A BAM made of thin material (30.0 μm) delivers high contraction of 37.0% while lifting 0.5 kg. (B) A BAM made of thicker material (125.0 μm) delivers high tensile force, lifting 3.0 kg at a contraction of 10.2%. BAMs, bubble artificial muscles.

The characteristics of the PPAM are based on actuator length L , actuator radius R , and applied pressure P , with the assumption of inelastic material behavior. The PPAM mathematical model was derived by using an elliptical integral with m and φ_R as dominant parameters to determine the actuator shape at any contraction. Three main equations are used to calculate the contraction c and tensile force T of the PPAM, as follows:

$$L = \frac{R}{\sqrt{m} \cos \varphi_R} F(\varphi_R/m) \quad (1)$$

$$c = 1 - \frac{2R}{L} \left(\frac{E(\varphi_R/m) - \frac{1}{2} F(\varphi_R/m)}{\sqrt{m} \cos \varphi_R} \right) \quad (2)$$

$$T = \pi P R^2 \frac{1 - 2m}{2m \cos^2 \varphi_R} \quad (3)$$

m defines the actuator's shape and its contraction, where zero contraction or maximum contraction are reached when m is equal to 0 or 0.5, respectively ($0 \leq m \leq 0.5$), as shown in Supplementary Figure S1B. φ_R is a characteristic angle mathematically related to the shape of the actuator membrane ($0 \leq \varphi_R \leq \pi/2$), calculated from Equation (1) when substituting an m value, and actuator size, L and R . This results in a φ_R value for each m value; $F(\varphi_R/m)$ and $E(\varphi_R/m)$ are the elliptic integrals of the first kind and second kind, respectively (Elliptic Integral section in Supplementary Data). c and T can then be calculated by Equations (2) and (3) respectively, given an applied pressure P . Full details related to the PPAM model can be found in the Pleated Pneumatic Artificial Muscles section in Supplementary Data.

Series pneumatic artificial muscle

The sPAM is made of a long plastic tubing and rubber O-rings, creating a series of pneumatic actuators similar to the PPAM.³⁸ Unlike the PPAM, an inactive region appears in the sPAM when the actuator length of the sPAM is greater than a certain value. This occurs when the actuator radius reaches the maximum material radius (the radius of the plastic tubing used to build the sPAM), causing the actuator to form a cylindrical region in the middle of the actuator, preventing the generation of further contraction, and thus limiting the overall contraction ratio. Although each contractile unit of the sPAM generates high contraction close to that of the PPAM model, it produces much lower tensile force than the PPAM due to its low material strength, which limits applied pressure.

Bubble artificial muscle

The BAM was developed to deliver high tensile force while maintaining high contraction. This is achieved by introducing stiff retaining rings and strong actuator material, allowing the BAM to sustain high pressure. The BAM is made by using a polyethylene plastic tubing and metal retaining rings (Fig. 2A). Two metal rings (gray in Fig. 2A) are inserted along the plastic tubing (pink in Fig. 2A) to create a single contractile unit. These metal rings constrain the tubing to form a tight folded shape within the ring radius (cross-section AA in Fig. 2A). Non-uniform folding of the plastic tubing extends along the actuator in the axial direction (red dashed lines in Fig. 2A). These induced folds function similarly to the pleats in the PPAM. The result is a PAM that is considerably less expensive and easier to manufacture compared with the PPAM, while exhibiting higher tensile force compared with the sPAM.

When inflating the BAM by pressurized air, the actuator membrane unfolds and expands radially, and thus the actuator contracts, forming different expanded shapes, that is, a vertical elliptical shape, a circular shape, and a horizontal elliptical shape, depending on the level of applied pressure P (Fig. 2B). Increasing the applied pressure results in an increase in actuator contraction. This increase happens only in the inelastic phase, whereby the membrane flexes and the actuator changes shape, with negligible elastic stretching of the membrane. If the pressure is increased beyond the point of maximum contraction, behavior enters an elastic phase.

In the inelastic phase (Fig. 2B), increasing the applied pressure causes the actuator membrane to stretch, inflating like a balloon, which results in a decrease in contraction. The optimal contraction happens when the actuator forms a horizontal elliptical shape in the inelastic phase, which resembles a “bubble,” leading to the name of the actuator, “Bubble Artificial Muscle.”

As shown in Figure 2A, the BAM is fabricated from a plastic tubing with a maximum material radius R_{material} and a material thickness s_{material} and metal rings with ring radius R_{ring} . With this design and fabrication method, the BAM can comprise many contractile units aligned in series as demonstrated in Figure 2C, enabling it to achieve any desired total strokes. For example, the BAM in Figure 2C consists of four contractile units; therefore, it can produce a maximum stroke four times that of a single unit. At the initial stage, the length of the entire BAM series of actuators is defined as an initial actuator length L_{actuator} , whereas that of each contractile unit is defined as an initial unit length L_{unit} . When inflated, the

BAM contracts by stroke ΔL and a total contraction c , where $c = \Delta L / L_{\text{actuator}}$. Its inflated unit length and inflated unit radius are defined as a bubble length L_{bubble} and a bubble radius R_{bubble} at maximum shape expansion ($m = 0.5$), respectively.

The optimal contraction c_{optimal} can be delivered by a BAM having L_{unit} equal to the optimal unit length L_{optimal} , actuated at the maximum applied pressure that retains inelastic behavior (resulting in maximum shape expansion). At any applied pressure P , a folded membrane around the metal rings always exists, forming an actuator shape as in cross-section AA in Figure 2C. However, when the optimal BAM is inflated, the amount of folded membrane is reduced incrementally with distance from the metal rings as can be seen in cross-sections AA, BB, and CC.

A fold length L_{fold} describes the length of the folded region of the BAM (red dashed lines) at maximum shape expansion in the axial direction, beginning from the actuator end and ending at the point where no folding appears on the membrane surface, shown as a black dashed line. Therefore, for the optimal BAM, L_{fold} is equal to 0.5 of L_{bubble} (or $L_{\text{bubble}} = 2L_{\text{fold}}$) as shown in Figure 2D (left). L_{fold} varies depended on the ring radius R_{ring} , material radius R_{material} , and material thickness s_{material} , causing a different amount of folding at the actuator ends.

Besides the optimal BAM, other expanded shapes can be obtained at the maximum applied pressure, depending on L_{unit} , R_{ring} , and R_{material} . For example, when an inactive region occurs, L_{inactive} defines the axial length of the inactive region (blue in Fig. 2D, middle), and L_{fold} is the same as that of the optimal BAM; therefore, $L_{\text{bubble}} = 2L_{\text{fold}} + L_{\text{inactive}}$. Alternatively, for shorter unit lengths, an overlapped region (red in Fig. 2D, right) occurs due to the crossover of folding from both actuator ends ($L_{\text{bubble, predicted}} < 2L_{\text{fold}}$). This results in an effectively stiffer actuator membrane, leading to partial shape expansion at the maximum applied pressure ($R_{\text{bubble, actual}} < R_{\text{material}}$) and lower contraction ($L_{\text{bubble, actual}} > L_{\text{bubble, predicted}}$).

The BAMs are designed to produce sufficiently high force and contraction to drive an orthotic to assist human muscles. As described earlier, BAM characteristics depend on an initial unit length L_{unit} , a ring radius R_{ring} , an applied pressure P , and a material thickness s_{material} . For example, the optimal “bubble” shape expansion and high contraction can be attained by carefully selecting L_{unit} and R_{ring} to avoid an overlapped region or an inactive region, whereas high force generation can be achieved by increasing s_{material} and maximizing P . Varying these four parameters leads to different performance of the BAM.

Experimental Setup

Several experiments were conducted to investigate the effects of P , L_{unit} , R_{ring} , and s_{material} to evaluate the resulting BAM performance. Three commercially available low-density polyethylene layflat tubes (Young’s Modulus $E = 0.3$ GPa) with different thicknesses ($s_{\text{material}} = 30.0, 62.5$ and $125.0 \mu\text{m}$) and similar radii (~ 17 mm) were selected for fabrication into BAMs as shown in Supplementary Table S1. The metal rings were made of metal with a thickness of 1.30 mm and an internal radius of 4.5 mm.

To evaluate the BAM performance, two types of experiments were undertaken: an isometric test and an isotonic test.

FIG. 2. (A) Components of a single-unit BAM: plastic tubing and two metal rings, and the cross-section of the BAM at the actuator end (AA). (B) BAMs at different applied pressure, demonstrating the inelastic and elastic phases. (C) Definitions of BAM parameters, images showing an optimal BAM at the initial stage and the inflated stage, and the cross-section of the actuator of these two stages at different points along the actuator (AA, BB, and CC). (D) Different expanded shapes of different single-unit BAMs at maximum applied pressure in the inelastic phase illustrating L_{bubble} and R_{bubble} at maximum shape expansion of (left) an optimal BAM, (middle) a BAM with an inactive region (blue area), and (right) a BAM with an overlapped region (red area). Color images are available online.

For isometric testing, the BAM was oriented vertically, with the top end mounted on an acrylic frame connected to a 1 kN load cell (700 Series S Beam Load Cell; Load Cell Shop, UK) and the bottom end attached to a linear actuator (LACT8P; Concentric International, USA). The tensile force was measured by the load cell through a load cell amplifier

(RW-ST01A; SMOWO, China), and the actuator stroke was controlled by the linear actuator and recorded by a laser displacement sensor (LK-G152; Keyence, Japan). Pressurized air was supplied by an air compressor (CW 100/24 AL; Werther International S.p.A., Germany) connected to the actuator through a pressure regulator (AR20-F02H010B;

SMC, UK) and a solenoid valve (WZ-98302-46; Cole-Parmer, UK), which were used to regulate the pressure level and inflate and deflate the actuator. A pressure sensor (HSCDANN030PGAA5; Honeywell, USA) was located close to the actuator to measure the applied pressure.

For isotonic testing, the bottom end of the actuator was disconnected from the linear actuator and a test mass was hung instead. The rest of the test environment was identical to the isometric test.

All tested BAMs consisted of four contractile units, excluding the optimal-unit-length experiment. Isometric tests were used to investigate the effects of varying P , L_{unit} , and R_{ring} of BAMs with the same $s_{material}$, and the relationship between tensile force and pressure, and between tensile force and contraction, of BAMs with different $s_{material}$. Isotonic tests were used to investigate the maximum contraction of BAMs while holding external loads. In the experimental results that follow, maximum contraction c_{max} is the contraction at zero tensile force ($T = 0$) and maximum tensile force T_{max} is the tensile force at zero stroke ($\Delta L = 0$).

Results

Increasing applied pressure (P)

A BAM with $L_{unit} = 40.5$ mm, $R_{ring} = 4.5$ mm, and $s_{material} = 62.5$ μm is considered here as an example to show the effect of increasing P . As shown in Figure 3A, the BAM achieves the maximum contraction of 32.0%, 34.2%, and 35.2% at $P = 10.0$, 20.0, and 30.0 kPa, respectively, with maximum contraction decreasing to 30.1% at $P = 40.0$ kPa.

This result can be explained by the progression from the inelastic phase to elastic phase, as presented in Figure 2B. Increasing applied pressure over a certain threshold (between 30 and 40 kPa in this case) causes the actuator membrane to stretch (behavior enters the elastic phase) and results in a reduction in the maximum contraction. Further increasing pressure would lead to irrecoverable plastic deformation and ultimately rupturing, therefore experiments were halted when elastic behavior occurred, before rupture. This stretching behavior can occur at any contraction when exceeding the pressure threshold (Supplementary Fig. S3A).

FIG. 3. (A) Maximum contraction (left y-axis) and maximum tensile force (right y-axis) of a BAM with $L_{unit} = 40.5$ mm, $R_{ring} = 4.5$ mm, and $s_{material} = 62.5$ μm , operated at different P . (B) Maximum contraction of a series of BAMs with different L_{unit} but constant $R_{ring} = 4.5$ mm and $s_{material} = 30.0$ μm , actuated at $P = 10.0$ kPa (the PPAM and sPAM models are plotted, and the image of the actual BAMs is inset). (C, D) Maximum contraction (left y-axis) and maximum tensile force (right y-axis) of the BAMs made of $s_{material} = 62.5$ μm with different slenderness ratios SR (L_{unit}/R_{ring}), where (C) $R_{ring} = 4.5$ mm and (D) $L_{unit} = 40.5$ mm, actuated at $P = 30.0$ kPa (images show the maximum expanded shape of the middle two contractile units of the tested BAMs). PPAM, pleated pneumatic artificial muscle; sPAM, series pneumatic artificial muscle. Color images are available online.

Varying initial unit length (L_{unit})

Five BAMs with L_{unit} of 31.5, 36.0, 40.5, 45.0, and 54.0 mm but constant R_{ring} of 4.5 mm (slenderness ratio $SR = L_{unit}/R_{ring} = 7, 8, 9, 10$ and 12) and $s_{material}$ of 62.5 μm were tested to demonstrate the effect of varying L_{unit} . Constant P of 30 kPa was used to inflate the BAMs without stretching behavior.

As shown in Figure 3C, the greatest maximum contraction occurs at $L_{unit} = 40.5$ mm ($SR = 9$). Lower maximum contraction appears at shorter L_{unit} ($SR = 7$ and 8) and longer L_{unit} ($SR = 10$ and 12) due to the appearance of the overlapped region and the inactive region, respectively. In contrast, the maximum tensile force increases when L_{unit} is higher. This is because the BAM has larger membrane surface area at $\Delta L = 0$ (membrane surface area is roughly equal to $2\pi R_{ring} L_{unit}$). The fluid pressure is applied over a larger surface, resulting in higher total radial force and thus higher tensile force. Further results at other stroke ranges can be found in Supplementary Figure S3.

Varying ring radius (R_{ring})

A BAM with constant L_{unit} of 40.5 mm and $s_{material}$ of 62.5 μm was created by using plastic cable ties in place of metal rings due to the ease in controlling internal radius. Ring radii of 6.750, 5.625, 4.500, 3.375, and 2.250 mm ($SR = 6, 7.2, 9, 12$, and 18) were selected to investigate the effect of varying R_{ring} . The actuators were inflated at constant $P = 30$ kPa.

From Figure 3D, decreasing R_{ring} results in higher maximum contraction and reduction of the inactive region. However, an overlapped region occurs when the BAM has too small R_{ring} , causing lower contraction than predicted by the PPAM model. As in the prior length-varying tests (Fig. 3C), the maximum tensile force increases with R_{ring} because of the larger membrane surface area at $\Delta L = 0$.

Optimal unit length ($L_{optimal}$)

The effects of the overlapped region and the inactive region, which influence $L_{optimal}$, were studied further with the BAM made of the thinnest material. A series of BAM units with different L_{unit} from 20.0 to 60.0 mm at increments of 5.0 mm were built as a single long actuator. Metal rings with

$R_{ring} = 4.5$ mm and a thin plastic tubing with $s_{material} = 30.0$ μm were selected for fabrication. $L_{optimal}$ was calculated as 35.2 mm (see BAM Optimal Unit Length and Optimal Ring Radius section in Supplementary Data). The experimental procedure involved inflating the BAM at $P = 10$ kPa, measuring its inflated length (using a calliper), and deflating. This experiment was repeated three times for each L_{unit} , and no external load was applied on the actuator. The maximum measured contraction of each single-unit BAM is presented in Figure 3B.

The largest contraction (mean at 36.9% and maximum at 38.7%) was observed for the BAM with $L_{unit} = 35.0$ mm, which is close to the predicted $L_{optimal}$ (35.2 mm). The BAMs with $L_{unit} < L_{optimal}$ follow the general trend of the PPAM model,³⁴ but there is an offset (a reduction in contraction) as a result of the overlapped membrane. The BAMs with $L_{unit} > L_{optimal}$ follow the general trend of the sPAM model.³⁸

Varying material thickness ($s_{material}$)

Isotonic and isometric tests were performed to evaluate the BAM's capability to produce contraction with an external load of 1 kg and maximum tensile force at $\Delta L = 0$ mm, respectively. The experimental results for BAMs with the same $L_{unit} = 40.5$ mm and $R_{ring} = 4.5$ mm but different $s_{material} = 30.0, 62.5$, and 125.0 μm are presented in Figure 4. The predicted PPAM contraction (which neglects the thickness of the actuator membrane) is included in Figure 4A: The PPAM model predicts infinite tensile force at zero contraction, so this is not included in Figure 4B.

Overall, at the same applied pressure, BAMs with thicker membranes produce less contraction and generally less tensile force, deviating increasingly from the PPAM model prediction (Fig. 4A). From Figure 4B, the tensile force of each BAM tends to decrease due to stretching behavior when applied pressure approaches their maximum pressure; increasing applied pressure further can cause the BAM to burst. The results of the isometric test measuring tensile force of the three BAMs at strokes other than zero are shown in Supplementary Figure S4.

Further isometric tests were performed to obtain the relationship between tensile force and contraction of BAMs with different $s_{material}$, as shown in Figure 5A. They were tested with different P of 10, 30, and 50 kPa, based on their $s_{material}$,

FIG. 4. Contraction and tensile force of BAMs ($L_{unit}=40.5$ mm and $R_{ring}=4.5$ mm) with three different $s_{material}$ undergoing (A) isotonic testing while with a 1 kg load and (B) isometric testing with $\Delta L=0$ mm; operated from 0 kPa to maximum pressures of 20, 40, and 60 kPa, respectively. The two vertical gray dashed lines indicate the maximum tested pressure of the BAMs made from 30 and 62.5 μm thick material, respectively. Color images are available online.

FIG. 5. (A) The relationship between tensile force and contraction for BAMs with different s_{material} but the same $L_{\text{unit}}=40.5 \text{ mm}$ and $R_{\text{ring}}=4.5 \text{ mm}$, operated with different applied pressure P . Dots indicate measured data from experiments, to which curve-fitting lines have been added. (B) Maximum contraction of BAMs made by using cable ties with $L_{\text{unit}}=40.5 \text{ mm}$, with different slenderness ratios SR ($L_{\text{unit}}/R_{\text{ring}}$), R_{ring} , and s_{material} , actuated at different P . PPAM model (dashed line) and sPAM model (dotted line) predictions are also shown. Color images are available online.

well below the maximum pressures observed in Figure 4. As a result, the thin-membrane BAM can produce the highest contraction of 39.5%, and the thick-membrane BAM can generate the highest tensile force up to 56.9 N.

Figure 5B shows the effect of decreasing R_{ring} on maximum contraction for BAMs made from three different s_{material} . BAMs made from $L_{\text{unit}}=40.5 \text{ mm}$ and $R_{\text{ring}}=6.750, 5.625, 4.500, 3.375$, and 2.250 mm were evaluated under zero tensile load. P was chosen to be well below the maximum pressure of each material. The theoretical maximum contractions of the PPAM and sPAM are also shown and are listed in Supplementary Table S2.

As can be seen in Figure 5B, discrepancy between the PPAM and sPAM models is maximized at high R_{ring} due to the inactive region. Decreasing R_{ring} reduces the size of the inactive region, and the sPAM model more closely matches the PPAM model. Decreasing R_{ring} increases the maximum contraction of the actual BAMs. However, it also results in a higher reduction in BAM contraction compared with the sPAM model due to the material thickness (both the PPAM and sPAM models assume a zero-thickness membrane). This behavior occurs with all s_{material} but is more pronounced with higher s_{material} . The sPAM model includes only the effect of the inactive region, but higher fold lengths, and a consequential overlapped region, are more likely to appear in actual BAMs with higher s_{material} . Consequently, to more accurately capture BAM actuation we must include these effects in the model.

Discussion

BAM performance

Typical pneumatic actuators can deliver a maximum contraction of around 25–35%, for example, McKibben muscle and Pouch Motor.^{26,36} Since the maximum expanded shape of the PPAM, sPAM, and BAM actuators is the horizontal ellipse (Fig. 2B), they can all feasibly reach a maximum contraction of 45.5% (Ref.³⁵). A single-contractile-unit PPAM with $L_{\text{actuator}}=100.0 \text{ mm}$, $R_{\text{actuator}}=12.5 \text{ mm}$ and weight of 58.3 g was able to deliver a maximum contraction of 41.5% and a maximum tension of 3500 N under $P=300 \text{ kPa}$.³⁴ An sPAM with $R_{\text{ring}}=2.5 \text{ mm}$ delivered a max-

imum contraction and tension of $\sim 40\%$ and 9 N, respectively, actuated at $P=10.34 \text{ kPa}$.³⁸

Compared with these PAMs, the BAM can be designed to deliver either high contraction when using the thinnest s_{material} and the smallest R_{ring} or high tension when using the thickest s_{material} and operating under high P . The highest contraction of 43.1% was delivered by the BAM with $s_{\text{material}}=30.0 \mu\text{m}$, $L_{\text{unit}}=40.5 \text{ mm}$, $R_{\text{ring}}=2.0 \text{ mm}$, and an actuator weight of 2.83 g while producing a maximum tensile force of 13.2 N.

Similarly, a BAM with $L_{\text{unit}}=40.5 \text{ mm}$, $R_{\text{ring}}=4.5 \text{ mm}$, and higher s_{material} of $125.0 \mu\text{m}$ delivered a maximum tensile force of 56.9 N ($=0.894 \text{ MPa}$), which corresponds to lifting a load 1000 times its own weight (5.39 g), and delivers a maximum contraction of 35.2% when operated at a pressure of $P=50.0 \text{ kPa}$ (Fig. 5A). Its weight is less than one tenth that of the PPAM because of its simpler actuator ends and lighter materials. Three BAMs made of the same actuator design and materials had a similar tension–contraction relationship; the sample variation is shown in Supplementary Figure S5.

Comparison between PPAM, sPAM, and BAM

Although the BAM superficially resembles the PPAM, they have different fundamental structures due to different actuator ends: a special end fitting for the PPAM and a metal ring for the BAM (Fig. 6A). The PPAM has uniform pleats in the radial direction, whereas the folds of the BAM are in the lateral direction. These lateral folds are freely and non-uniformly folded around the metal ring, creating a region of overlapping folds, which naturally resists the radial expansion of the BAM. As membrane thickness increases, the amount of lateral folds also increases (Fig. 6B), increasing resistance to BAM shape expansion and reducing contraction and tension. However, higher thickness materials can withstand higher applied pressures, leading to higher expanding force and thus higher tensile force and maximum contraction of the BAM when loaded (Fig. 4).

When the BAM is inflated, the folds unfold by sliding and bending to expand the actuator shape (Fig. 6A), approaching the circular actuator cross-section (CC in Fig. 2C). The thickness of the folded membrane along the actuator axis x

FIG. 6. (A) Cross-section view of the actuator end, side view, cross-section lateral view, and cross-section axial view showing membrane expansion, comparing the PPAM and BAM. The *pink area* is the actuator's material, and the *gray area* is the actuator ends: the special end fitting for the PPAM and the metal ring for the BAM. The *pink color* indicates thin membrane, whereas the *red color* indicates where multiple membrane folds occur. (B) Actuator side and end view of BAMs with different $s_{material}$ of 30.0, 62.5, and 125.0 μm , but the same $L_{unit}=40.5$ mm and $R_{ring}=4.5$ mm. As material thickness is increased, the increase in fold length is clearly visible. Color images are available online.

may be defined as the bubble surface thickness $s_{bubble}(x_i)$, where x_i is the distance from the actuator center (Supplementary Fig. S1). $s_{bubble}(x_0)$ is highest at the actuator ends (higher than $s_{material}$), where many overlapping folds occur, and it reduces toward the middle of the actuator where no overlapping folds occur such that $s_{bubble}(x_n) = s_{material}$ when inflated (Fig. 6A). Unlike the BAM, the PPAM has no overlapping folds due to its radial folding structure, resulting in zero-friction shape expansion when inflated.

Although the sPAM and BAM share some functionality, the sPAM was designed to control the movement of light-weight continuum robots. In contrast, the BAM was designed

as a high-power artificial muscle (e.g., for wearable assistive applications), necessitating higher contraction and tension. Various aspects of the BAM design result from these higher required performance metrics (high-thickness plastic tubing that allows for considerably higher applied pressure, metal retaining rings that withstand high radial force and folding analysis).

An isotonic test using a load of 1 kg demonstrates the difference in performance between these two actuators (Supplementary Fig. S6). With the BAM, the metal rings maintained an effective actuator shape, resulting in the BAM contraction after theoretical PPAM contraction. In contrast,

the rubber rings of the sPAM stretched as applied pressure was increased, inducing an inactive region and causing large deviation from the PPAM model. Maximum BAM contraction was 26.5% at 40 kPa, significantly larger than the sPAM contraction of 18.2% at the same pressure.

BAM characterization

The BAM introduces new behavior, an overlapped region, which, together with the inactive region present in the sPAM (Fig. 2D), reduces maximum contraction and causes deviation from the PPAM mathematical model. The overlapped region occurs when there is a large amount of folds around the actuator ends, causing crossover of the folds from two ends and overlapped membrane across the actuator. It increases the membrane stiffness and difficulty in unfolding and bending of the folded membrane, and it reduces the bubble radius R_{bubble} . This stops the actuator from reaching its optimum shape, leading to partial shape expansion and lower maximum contraction.

When fixing R_{ring} and varying L_{unit} (Fig. 7A), the optimal “bubble” shape with $c_{optimal}$ can be achieved when the BAM possesses $L_{unit} = L_{optimal}$ as shown in Figure 7A, column A2. Decreasing L_{unit} below $L_{optimal}$ ($L_{unit} < L_{optimal}$) causes a partial shape expansion, creating overlapping of the material on the actuator surface, shown as the red area in Figure 7A, column A1. On the other hand, increasing L_{unit} over $L_{optimal}$ ($L_{unit} > L_{optimal}$) causes an inactive region, shown as the blue area in Figure 7A, column A3.

Likewise, the overlapped region and the inactive region occur when varying R_{ring} (Fig. 7B). When fixing L_{unit} and increasing R_{ring} , the radius of the actuator ends approaches the maximum material radius $R_{material}$, reducing the amount of folding at the actuator ends and the fold length L_{fold} until $L_{fold} = 0$ at $R_{ring} = R_{material}$, as shown in columns B1, B2, B3, and B4, respectively (the black dashed line shows the end of the folds). This leads to the inactive region shown in Figure 7B,

column B3 and eventually no contraction in column B4. L_{fold} is unaffected by L_{unit} (Fig. 7A, columns A1, A2, and A3).

Although the inactive region can be addressed by decreasing R_{ring} , too small R_{ring} can result in the overlapped region as shown in Figure 7B, column B1. The optimal ring radius $R_{optimal}$ will create neither an inactive region nor an overlapped region, and it will deliver the highest contraction (Fig. 7B, column B2). The approach to calculate $L_{optimal}$ and $R_{optimal}$ to design the optimal BAM, which produces $c_{optimal}$, is presented in the BAM Optimal Unit Length and Optimal Ring Radius section in Supplementary Data.

BAM actuation model

Although the BAM produces lower performance than the PPAM theoretical maximum (Figs. 4 and 5) because of the fundamental difference in their folding structure, they share the same ideal behavior of shape expansion (Fig. 2B). Consequently, the BAM mathematical model is built on the PPAM model with the addition of the inactive region modeled by the sPAM model, and modifications to model the effects of material thickness (Supplementary Figs. S7 and S8). The PPAM and sPAM models overestimate BAM performance in terms of contraction and tensile force, as they do not account for material thickness, a major factor in BAM actuation. This deviation can be reduced by modifying the inelastic PPAM model with an additional term based on experimental observation. The effect of the inactive region from the sPAM model, which can limit the maximum contraction, is also included in the BAM model.

As described earlier in the Varying Material Thickness ($s_{material}$) section and Figure 5B, the ratio of the material thickness and the ring radius ($s_{material}/R_{ring}$) and the ratio of the material radius and the ring radius ($R_{material}/R_{ring}$) are likely to be dominant factors in determining the reduction in the BAM performance compared with the PPAM model. As

FIG. 7. The BAM parameter (top row), the predicted PPAM shape (middle row), and the actual BAM shape (bottom row) at maximum shape expansion when (A) varying L_{unit} with constant R_{ring} and (B) varying R_{ring} with constant L_{unit} . Color images are available online.

such, the loss in contraction c and tensile force T can be described by the following equation, where A and n are constant.

$$loss \propto A * \left(\frac{s_{material}}{R_{ring}} * \frac{R_{material}}{R_{ring}} \right)^n \quad (4)$$

The relationship between tensile force and contraction of the BAM (Fig. 5A) can be modified from the PPAM model by using the subtractive contraction loss c_{loss} , applying from Equation (4). Therefore, the BAM contraction c_{BAM} can be derived from Equation (5), where c_{PPAM} is the PPAM contraction, and where A and n have been chosen to fit experimental data.

$$\begin{aligned} c_{BAM} &= c_{PPAM} - c_{loss} \text{ when } c_{loss} \\ &= 21.35 * \left(\frac{s_{material} R_{material}}{R_{ring}^2} \right)^{0.25} \end{aligned} \quad (5)$$

The comparison between the PPAM model, the sPAM model, the modified model for the BAM, and the experi-

mental data of the BAM ($s_{material} = 62.5 \mu\text{m}$ from Fig. 5A) is presented in Figure 8A. Applying c_{loss} from Equation (5) causes shifting of the PPAM model to match the BAM experimental data (Supplementary Fig. S7). Although not required for the results in Figure 8A, the BAM model also includes the limitation on maximum contraction due to the inactive region, first described in the sPAM model (Supplementary Figs. S7 and S8). The BAM model for the BAMs made of different $s_{material}$ (Fig. 5A) can be seen in Figure 8B. This model can also be used to predict the real-world contraction of the BAM with constant $s_{material}$ at various applied pressures (Fig. 8C).

In contrast, the relationship between contraction and pressure (Fig. 4A) is better modeled by applying a multiplicative tension loss, $T_{loss} = 1/\eta_{loss}$ to PPAM tension T_{PPAM} as in Equation (6). As previously shown experimentally, the BAM requires higher applied pressure P than predicted by the PPAM model to deliver desired T and c . Conversely, this means that c_{BAM} is less than c_{PPAM} at the same P and T . Figure 8D illustrates the increasing divergences of c_{BAM} from c_{PPAM} with $s_{material}$ while loaded with $T = 1 \text{ kg}$.

FIG. 8. (A) Comparison between the PPAM model, sPAM model, and BAM model for a BAM with $L_{unit}=40.5 \text{ mm}$, $R_{ring}=4.5 \text{ mm}$, and $s_{material}=62.5 \mu\text{m}$, operated at $P=30 \text{ kPa}$, showing the relationship between tensile force and contraction. The maximum sPAM contraction without an inactive region occurring, and the BAM experimental data are included. (B, C) BAM model applying subtractive contraction loss (lines) and experimental results (dots) of BAMs with $L_{unit}=40.5 \text{ mm}$ and $R_{ring}=4.5 \text{ mm}$, showing the relationships between tensile force and contraction when (B) varying $s_{material}$ and P and (C) having constant $s_{material}=62.5 \mu\text{m}$ and operating at different P . (D) BAM model applying multiplicative tension loss (lines) and experimental results (dots) of BAMs with $L_{unit}=40.5 \text{ mm}$ and $R_{ring}=4.5 \text{ mm}$, showing the relationship between contraction and pressure when varying $s_{material}$ and operating at different P . Color images are available online.

$$T_{BAM} = \frac{1}{\eta_{loss}} * T_{PPAM} \text{ when } \eta_{loss} = 4.39 * \left(\frac{s_{material} R_{material}}{R_{ring}^2} \right)^{0.31} \quad (6)$$

Design optimization

Analysis summarized in Figure 9 shows how the BAM actuator differs from the PPAM and the sPAM, and how it can be optimized to achieve high contraction by choosing three ratios of material properties: L_{unit}/R_{ring} , $R_{ring}/R_{material}$, and $s_{material}/R_{ring}$. Figure 9A shows the maximum contraction predicted by the PPAM model, simulated by substituting L_{unit}/R_{ring} in Equations (1) and (2). It does not show an overlapped region or an inactive region since the PPAM actuator membrane is uniformly folded and its actuation never reaches maximum shape expansion, and its contraction is maximized as L_{unit}/R_{ring} and does not depend on $R_{ring}/R_{material}$.

Figure 9B shows the maximum contraction predicted by the sPAM model. At any $R_{ring}/R_{material}$, contraction increases with L_{unit}/R_{ring} until reaching a maximum when $L_{unit} = L_{optimal}$. If L_{unit}/R_{ring} is further increased, contraction reduces due to the inactive region. Neither the PPAM nor the sPAM models account for the effect of $s_{material}$, which is a significant factor influencing the contractile performance. Figure 9C shows the BAM model presented in this article; the effect of material thickness slightly reduces maximum contraction, and it slightly increases the optimal L_{unit}/R_{ring} at each $R_{ring}/R_{material}$ compared with the sPAM model. The optimal design of the BAM made of different $s_{material}$ is shown in Supplementary Figure S9.

BAM improvement

The performance of the BAM can be improved by several approaches. First, better quality fabrication and a method for creating uniformly lateral folding along the actuator are required so that the BAM can unfold and expand more easily with lower friction. Uniform folding can increase BAM contraction and decrease the deviation between the mathematical model and the practical performance.

For the current design, the rings are placed without any attachments to the tubing. After repeated actuation, the rings tend to be fixed in place by the shape adopted by the folded membrane; as $R_{material}$ is much larger than R_{ring} , the folded membrane passively bends around the ring, helping to constrain the movement of the rings. If BAM actuators are to be integrated into a robotic system without prior actuation, the rings can be fixed in place by adhesive to prevent slippage. Alternatively, one future BAM design could have no rings but would use its own actuator membrane to form the “bubble” shape, for example, by applying origami or kirigami methods so that the bubble shape naturally emerges.

Last, using high-strength inelastic materials, the BAM can operate under higher applied pressure to deliver higher tensile force. The ideal actuator membrane should possess high Young's Modulus and extremely high tensile strength to withstand high pressure for high tensile force, have low interfacial friction, and be very thin and very flexible to expand easily at low-pressure actuation and deliver high contraction.

BAM Orthosis for Sit-to-Stand Transition

A leg mechanism was built to evaluate the BAM performance in assisting human mobility in the task of standing up from a sitting position (sit-to-stand), as shown in Figure 10A. It consists of three segments: a foot base, a shank rod, and a thigh rod, which are connected via revolute joints representing the knee and ankle (Fig. 10B). It was designed so that the shank rod can be fixed at an adjustable ankle angle and an external load can be added to the hip joint, to represent a body weight. A BAM orthosis was created by using three pairs of BAMs aligned in parallel (Fig. 10D). In each pair, BAMs are connected in series by cables and located at the thigh and shank at 7.5 cm from the center axis, labeled as BAM₁ and BAM₂, respectively (Fig. 10C).

Each BAM contains seven contractile units, where $L_{actuator} = 305.0$ mm, $L_{unit} = 40.5$ mm, $R_{ring} = 4.5$ mm, and $s_{material} = 125.0$ μ m. The ends of BAM₁ and BAM₂ connect to the hip and ankle anchors and are joined by connecting cables that pass through the knee anchor. Based on the maximum tensile force of a single BAM of 56.9 N, the effective support force for the sit-to-stand transition was 171 N. The BAM orthosis was able to perform sit-to-stand transition within 5 s, supplied by the air compressor at $P = 50$ kPa (Fig. 10D). The

FIG. 9. Model-predicted maximum contraction of (A) PPAM, (B) sPAM, and (C) BAM made of $s_{material} = 30.0$ μ m. Color shows the maximum contraction c_{max} of the actuators as L_{unit}/R_{ring} and $R_{ring}/R_{material}$ are varied. For each $R_{ring}/R_{material}$, dashed and solid lines indicate the location where maximum contraction is highest based on the sPAM and BAM model, respectively. Color images are available online.

FIG. 10. The experimental setup of the sit-to-stand motion (A) is composed of two main parts: the leg mechanism (B) and the BAM orthosis (C). (D) Actuation of the BAM orthosis on the human-like leg mechanism to perform standing motion, operated at 50 kPa. Color images are available online.

device could be attached to a human subject by using tight straps, or it could be attached to a passive orthotic such as a knee-brace to convert it to an assistive device.

Conclusion

The BAM is one of the most lightweight pneumatic actuators (<6 g) that can deliver either high contraction or high tension depending on the actuator's size and thickness and the stiffness of the constituent material. A thicker actuator

membrane allows the BAM to operate under higher applied pressure to produce higher tensile force, whereas metal rings can maintain the actuator radius, delivering desired contraction when operated at high pressure. Despite having a different folding pattern to the PPAM, the BAM can form a similar inflated shape and function but with considerably simpler and low-cost fabrication. The optimal BAM can be achieved by choosing the unit length and ring radius to suit the material radius and thickness, avoiding both an overlapped region and an inactive region.

The BAM actuation model was built based on the PPAM model with an additional loss term, modifying the model to match the BAM experimental results. This model can predict the real-world performance for BAMs made from different material thicknesses and operated at different applied pressures. In this work, the loss term was empirically derived, and it demonstrates how BAM and sPAM behavior differ and how BAM performance can be improved. In the future, we plan to derive a BAM loss term from first principles and compare this theoretical model with the empirically derived model.

The BAM was designed to interact with the human body in the form of a wearable exosuit or orthosis. It fulfills this design brief by being lightweight and scalable, and by generating human-scale forces and contractions at low pressure. An example of sit-to-stand transition was demonstrated by using a leg mechanism assisted by a BAM orthosis, exhibiting sit-to-stand within 5 s. The BAM can be improved further by exploiting actuator materials that possess high strength and flexibility to achieve higher tension and contraction, removing the requirement of retaining rings by using its own material as the structure to form a contractile shape, and improving fabrication quality to create uniform folding for lower-friction shape expansion.

Data Access Statement

Data necessary to support the conclusions are included in the article.

Author Disclosure Statement

No competing financial interests exist.

Funding Information

This work is supported by the EPSRC grant EP/L015293/1 (Centre for Doctoral Training in Future Autonomous and Robotic Systems—FARSCOPE) at Bristol Robotics Laboratory where R.S.D. is a PhD student. T.H. is supported by EPSRC grant EP/M026388/1 and by the Royal Academy of Engineering and the Office of the Chief Science Adviser for National Security under the UK Intelligence Community Postdoctoral Fellowship Programme. M.T. is supported by EPSRC grants EP/M026388/1 and EP/R02961X/1. J.R. is supported by EPSRC grants EP/M026388/1, EP/M020460/1, EP/S026096/1, EP/S021795/1, and EP/R02961X/1 and by the Royal Academy of Engineering as Chair in Emerging Technologies.

Supplementary Material

Supplementary Data
 Supplementary Figure S1
 Supplementary Figure S2
 Supplementary Figure S3
 Supplementary Figure S4
 Supplementary Figure S5
 Supplementary Figure S6
 Supplementary Figure S7
 Supplementary Figure S8
 Supplementary Figure S9
 Supplementary Table S1
 Supplementary Table S2
 Supplementary Movie S1

References

1. Yan T, Cempini M, Oddo CM, *et al.* Review of assistive strategies in powered lower-limb orthoses and exoskeletons. *Rob Auton Syst* 2015;64:120–136.
2. Rupal BS, Rafique S, Singla A, *et al.* Lower-limb exoskeletons: research trends and regulatory guidelines in medical and non-medical applications. *Int J Adv Robot Syst* 2017;14:1729881417743554.
3. Dzahir M, Yamamoto S. Recent trends in lower-limb robotic rehabilitation orthosis: control scheme and strategy for pneumatic muscle actuated gait trainers. *Robotics* 2014; 3:120–148.
4. Sawicki GS, Ferris DP. A pneumatically powered knee-ankle-foot orthosis (KAFO) with myoelectric activation and inhibition. *J Neuroeng. Rehabil* 2009;6:23.
5. Wilkening A, Hacker S, Stöppler H, *et al.* Experimental and simulation-based investigation of polycentric motion of an inherent compliant pneumatic bending actuator with skewed rotary elastic chambers. *Robotics* 2017;6:2.
6. Park Y-L, Chen B, Young D, *et al.* Bio-inspired active soft orthotic device for ankle foot pathologies. In: 2011 IEEE/RSJ International Conference on Intelligent Robots and Systems, San Francisco, CA, September 25–30, 2011, pp. 4488–4495. IEEE.
7. Park Y-L, Chen B, Majidi C, *et al.* Active modular elastomer sleeve for soft wearable assistance robots. In: 2012 IEEE/RSJ International Conference on Intelligent Robots and Systems, Vilamoura, Portugal, October 7–12, 2012, pp. 1595–1602. IEEE.
8. Wehner M, Quinlivan B, Aubin PM, *et al.* A lightweight soft exosuit for gait assistance. In: 2013 IEEE International Conference on Robotics and Automation, Karlsruhe, Germany, May 6–10, 2013, pp. 3362–3369. IEEE.
9. Kurumaya S, Suzumori K, Nabae H, *et al.* Musculoskeletal lower-limb robot driven by multifilament muscles. *Robomech J* 2016;3:18.
10. Abe T, Koizumi S, Nabae H, *et al.* Muscle textile to implement soft suit to shift balancing posture of the body. In: 2018 IEEE International Conference on Soft Robotics (RoboSoft), Livorno, Italy, April 24–28, 2018, pp. 572–578. IEEE.
11. Case JC, Booth J, Shah DS, *et al.* State and stiffness estimation using robotic fabrics. In: 2018 IEEE International Conference on Soft Robotics (RoboSoft), Livorno, Italy, April 24–28, 2018, pp. 522–527. IEEE.
12. Chang S-Y, Takashima K, Nishikawa S, *et al.* Design of small-size pouch motors for rat gait rehabilitation device. In: 2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), Milan, Italy, August 25–29, 2015, pp. 4578–4581. IEEE.
13. Sridar S, Qiao Z, Muthukrishnan N, *et al.* A soft-inflatable exosuit for knee rehabilitation: assisting swing phase during walking. *Front Robot AI* 2018;5:44.
14. Kawamura T, Takanaka K, Nakamura T, *et al.* Development of an orthosis for walking assistance using pneumatic artificial muscle: a quantitative assessment of the effect of assistance. In: 2013 IEEE 13th International Conference on Rehabilitation Robotics (ICORR), Seattle, WA, June 24–26, 2013, pp. 1–6. IEEE.
15. Park Y-L, Santos J, Galloway KG, *et al.* A soft wearable robotic device for active knee motions using flat pneumatic artificial muscles. In: 2014 IEEE International Conference on Robotics and Automation (ICRA), Hong Kong, China, May 31–June 7, 2014, pp. 4805–4810. IEEE.

16. Asbeck AT, Dyer RJ, Larusson AF, *et al.* Biologically-inspired soft exosuit. In: 2013 IEEE 13th International Conference on Rehabilitation Robotics (ICORR), Seattle, WA, June 24–26, 2013; pp. 1–8. IEEE.
17. Schmidt K, Duarte JE, Grimmer M, *et al.* The Myosuit: bi-articular anti-gravity exosuit that reduces hip extensor activity in sitting transfers. *Front Neurobot* 2017;11:57.
18. Poliero T, Di Natali C, Sposito M, *et al.* Soft wearable device for lower limb assistance: assessment of an optimized energy efficient actuation prototype. In: 2018 IEEE International Conference on Soft Robotics (RoboSoft), Livorno, Italy, April 24–28, 2018, pp. 559–564. IEEE.
19. Li Y, Hashimoto M. Design and prototyping of a novel lightweight walking assist wear using PVC gel soft actuators. *Sensors Actuators A Phys* 2016;239:26–44.
20. Li Y, Maeda Y, Hashimoto M. Lightweight, soft variable stiffness gel spats for walking assistance. *Int J Adv Robot Syst* 2015;12:1.
21. Kovacs G, Düring L, Michel S, *et al.* Stacked dielectric elastomer actuator for tensile force transmission. *Sensors actuators A Phys* 2009;155:299–307.
22. Anderson IA, Gisby TA, McKay TG, *et al.* Multifunctional dielectric elastomer artificial muscles for soft and smart machines. *J Appl Phys* 2012;112:41101.
23. Taghavi M, Helps T, Rossiter J. Electro-ribbon actuators and electro-origami robots. *Sci Robot* 2018;3:eaau9795.
24. Kellaris N, Venkata VG, Smith GM, *et al.* Peano-HASEL actuators: muscle-mimetic, electrohydraulic transducers that linearly contract on activation. *Sci Robot* 2018;3:eaar3276.
25. Haines CS, Lima MD, Li N, *et al.* Artificial muscles from fishing line and sewing thread. *Science* (80-) 2014;343:868–872.
26. Daerden F, Lefeber D. Pneumatic artificial muscles: actuators for robotics and automation. *Eur J Mech Environ Eng* 2002;47:11–21.
27. Tondu B, Lopez P. Modeling and control of McKibben artificial muscle robot actuators. *IEEE Control Syst Mag* 2000;20:15–38.
28. Carlo Ferraresi WF, Walter Franco W, Bertetto A. Flexible pneumatic actuators: a comparison between the McKibben and the straight fibres muscles. *J RobMechatronics* 2001;13:56–63.
29. Nakamura T, Saga N, Yaegashi K. Development of a pneumatic artificial muscle based on biomechanical characteristics. In: IEEE International Conference on Industrial Technology, 2003, Maribor, Slovenia, December 10–12, 2003, Vol. 2, pp. 729–734. IEEE.
30. Tomori H, Nakamura T. Theoretical comparison of McKibben-type artificial muscle and novel straight-fiber-type artificial muscle. *Int J Autom Technol* 2011;5:544–550.
31. Durante F, Antonelli MG, Zobel PB, *et al.* Development of a straight fibers pneumatic muscle. *Int J Autom Technol* 2018;12:413–423.
32. Martinez RV, Fish CR, Chen X, *et al.* Elastomeric origami: programmable paper-elastomer composites as pneumatic actuators. *Adv Funct Mater* 2012;22:1376–1384.
33. Schaffner M, Faber JA, Pianegonda L, *et al.* 3D printing of robotic soft actuators with programmable bioinspired architectures. *Nat Commun* 2018;9:878.
34. Daerden F. Conception and realization of pleated pneumatic artificial muscles and their use as compliant actuation elements. PhD thesis, Vrije Universiteit Brussel, Brussels, 1999;176.
35. Daerden F, Lefeber D. The concept and design of pleated pneumatic artificial muscles. *Int J Fluid Power* 2001;2:41–50.
36. Niiyama R, Rus D, Kim S. Pouch motors: printable/inflatable soft actuators for robotics. In: 2014 IEEE International Conference on Robotics and Automation (ICRA), Hong Kong, China, May 31–June 7, 2014, pp. 6332–6337. IEEE.
37. Felt W. Folded-tube soft pneumatic actuators for bending. *Soft Robot* 2019;6:174–183.
38. Greer JD, Morimoto TK, Okamura AM, *et al.* Series pneumatic artificial muscles (SPAMs) and application to a soft continuum robot. In: 2017 IEEE International Conference on Robotics and Automation (ICRA), Singapore, Singapore, May 19–June 3, 2017, pp. 5503–5510. IEEE.
39. Greer JD, Morimoto TK, Okamura AM, *et al.* A soft, steerable continuum robot that grows via tip extension. *Soft Robot* 2019;6:95–108.
40. Diteesawat RS, Helps T, Taghavi M, *et al.* High strength bubble artificial muscles for walking assistance. In: 2018 IEEE International Conference on Soft Robotics (RoboSoft), Livorno, Italy, April 24–28, 2018, pp. 388–393. IEEE.

Address correspondence to:

Jonathan Rossiter
 Department of Engineering Mathematics
 University of Bristol
 Bristol BS8 1UB
 United Kingdom

E-mail: jonathan.rossiter@bristol.ac.uk